

**Bid Bulletin No. 1**  
May 29, 2019

**Public Bidding 19-160-3**

**Supply and Delivery of Eye Screening Equipment  
for the Department of Health (DOH)**

Issued pursuant to Sec. 22.5 of the IRR of R.A. 9184 to clarify, include and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by prospective bidders during the **Pre-Bid Conference** held on **15 May 2019**, likewise, respond to bidders' written queries received within the prescriptive period for filing.

**A. AMENDMENT**

REFERENCE	BASES FOR CLARIFICATION/ AMENDMENT/INCLUSION								
<b>SECTION I. INVITATION TO BID</b>									
<p><i>page 4</i></p> <p><i>xxx</i></p> <p><i>1. The summary of the bidding activities is as follows:</i></p> <table border="1"> <thead> <tr> <th align="center">Activities</th> <th align="center">Schedule</th> </tr> </thead> <tbody> <tr> <td align="center">xxx</td> <td align="center">xxx</td> </tr> <tr> <td>Deadline of Submission of Bids</td> <td align="center"><del>June 5, 2019, 1:30PM</del> <b>June 6, 2019, 3:00PM</b></td> </tr> <tr> <td>Opening of Bids</td> <td align="center"><b>Immediately after the Deadline of Submission of Bids</b></td> </tr> </tbody> </table> <p><i>xxx</i></p>	Activities	Schedule	xxx	xxx	Deadline of Submission of Bids	<del>June 5, 2019, 1:30PM</del> <b>June 6, 2019, 3:00PM</b>	Opening of Bids	<b>Immediately after the Deadline of Submission of Bids</b>	<p>To amend the Schedule of Deadline of Submission of Bids.</p>
Activities	Schedule								
xxx	xxx								
Deadline of Submission of Bids	<del>June 5, 2019, 1:30PM</del> <b>June 6, 2019, 3:00PM</b>								
Opening of Bids	<b>Immediately after the Deadline of Submission of Bids</b>								

*For the purpose of this Bulletin and for better understanding of its contents, the following rules shall apply: (a) ~~double~~ ~~Strikeout~~ – denotes deletion; (b) Underline – denotes inclusion or new item/requirement; and “xxx” – denotes separation of phase/s being amended from the rest of the main text.*

**SECTION III. BID DATA SHEET**

*ITB Clause 29.2*

*page 43*

xxx

6. The bidder shall submit any of the following **documents** whichever is applicable:

xxx

To clarify the item requirement for better understanding, see attached Appendix "A"

Section III. Bid Data Sheet

**SECTION VII. TECHNICAL SPECIFICATIONS**

*page 73*

LOT NO. 1	:	<b>Supply and Delivery of Eye Screening Equipment</b>
ITEM NO.1	:	<b>Snellen Chart, Lighted</b>
QUANTITY	:	1,419 <del>sets</del> <b>units</b>

To amend the unit of measure, see attached Appendix "B"

Section VII. Technical Specifications

<b>AGENCY SPECIFICATIONS</b>	<b>BIDDER'S STATEMENT OF COMPLIANCE</b>
<b>Snellen Chart, Lighted</b>	Brand and model:
xxx	
Thickness: 8 inches ( <b>maximum</b> )	
xxx	
With <b>maximum of</b> five (5) bulbs/ <del>tubes</del>	
<b><u>With minimum total power consumption of 25 watts</u></b>	
<b><u>xxx</u></b>	
<b><u>Printing on the Snellen Chart will be in Letters</u></b>	
<b><u>Can be read at a distance of 10 feet</u></b>	
xxx	
REQUIRED DOCUMENTS:	

To amend the form, see attached Appendix "B"

Section VII. Technical Specifications

1. The bidder shall submit any of the following **documents** whichever is applicable:

xxx

page 75

LOT NO. 1	:	<b>Supply and Delivery of Eye Screening Equipment</b>
ITEM NO.2	:	<b>Eye Occluder, Plastic</b>
QUANTITY	:	3,000 sets

To amend the form, see attached Appendix "B" Section VII. Technical Specifications

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE
<b>Eye Occluder, Plastic</b>	Brand and model:
xxx	
REQUIRED DOCUMENTS: 1. The bidder shall submit any of the following <b>documents</b> whichever is applicable: xxx	

page 77

LOT NO.1	:	<b>Supply and Delivery of Eye Screening Equipment</b>
ITEM NO.3	:	<b>Pinhole Eye Occluder, Plastic</b>
QUANTITY	:	3,000 units

To amend the form, see attached Appendix "B" Section VII. Technical Specifications

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE

<b>Pinhole Eye Occluder, Plastic</b>	Brand and model:
xxx	
<b>REQUIRED DOCUMENTS:</b>  1. The bidder shall submit any of the following <b><u>documents</u></b> whichever is applicable:  xxx	

page 79

LOT NO.1	:	<b>Supply and Delivery of Eye Screening Equipment</b>
ITEM NO.4	:	<b>Diagnostic Set with Ophthalmoscope, Fiber-Optic Oscope</b>
QUANTITY	:	2,100 sets

To amend the form, see attached Appendix "B" Section VII. Technical Specifications

<b>AGENCY SPECIFICATIONS</b>	<b>BIDDER'S STATEMENT OF COMPLIANCE</b>
<b>Diagnostic Set with Ophthalmoscope, Fiber-Optic Oscope</b>	Brand and model:
<b>Ophthalmoscope</b> <ul style="list-style-type: none"> <li>• Pocket type</li> <li>• Ophthalmoscope</li> <li>• Illumination/Direct Illumination: 18.5 Lumen (<b><u>minimum</u></b>)</li> <li>• LED</li> <li>• 2.5 volts (<b><u>minimum</u></b>)</li> <li>• <b><u>One (1) piece (head and handle)</u></b></li> </ul>	
<b>Fiber-Optic Oscope</b> <ul style="list-style-type: none"> <li>• High Performance Fibre Optics with 4mm diameter</li> <li>• 2.5V LED <b><u>or</u></b> Xenon &amp; Vacuum Illuminations</li> <li>• <b><u>2 One (1) pieces</u></b> (<del>1 piece</del> head and <del>1 piece</del> handle)</li> <li>• 3x magnifying lens and wide viewing window</li> <li>• Connection via screw thread and</li> </ul>	

<ul style="list-style-type: none"> <li>• dimmable light control system</li> <li>• <del>14</del> <b>Five (5)</b> Disposable ear funnels <u>per</u> size: 2.5 mm, 3.0mm, 4.0mm</li> </ul>		
<b>Accessories:</b> <ul style="list-style-type: none"> <li>• Hard Storage</li> <li>• Accessory bumper kit</li> <li>• Onyx <b>color</b></li> <li>• Battery included (AA batteries)</li> </ul>		
Warranty: Must have at least five (5) years warranty upon delivery of goods. <u><b>Warranty of the unit does not include the bulb.</b></u>		
<b>REQUIRED DOCUMENTS:</b> <ol style="list-style-type: none"> <li>1. The bidder shall submit any of the following <b>documents</b> whichever is applicable:</li> </ol> xxx		

**B. CLARIFICATIONS**

<b>REFERENCE</b>	<b>CLARIFICATION</b>	<b>RESOLUTION</b>
Section VII. Technical Specifications Page 75	Item No. 2: Eye Occluder, Plastic  Accessories: Carrying Case with zipper to be deleted	Item No. 2: Eye Occluder, Plastic  Retain the Carrying Case with zipper
Section VII. Technical Specifications Page 77	Item No. 3: Pinhole Eye Occluder, Plastic  Accessories: Carrying Case with zipper to be deleted	Item No. 3: Pinhole Eye Occluder, Plastic  Retain the Carrying Case with zipper
Section VII. Technical Specifications Page 79	Item No. 4: Diagnostic Set with Ophthalmoscope, Fiber-Optic Otoscope  Accessories: Hard storage, include also the Soft storage	Item No. 4: Diagnostic Set with Ophthalmoscope, Fiber-Optic Otoscope  Retain the Hard storage
	Country of Origin: Europe	No need to indicate the Country of Origin because it will limit the bidding.

All other portions of the Bidding Documents affected by these amendments shall be made to conform to the same.

Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

The changes made in the Philippine Bidding Documents (5<sup>th</sup> Edition, August 2016) are deemed integrated in the terms and conditions for this project.

SGD  
**MARC DLR. HORDEJAN**  
Vice-Chairperson, DBM-PS BAC 3

## Section III. Bid Data Sheet

29.2	<p>No other licenses and permits required.</p> <p>For purposes of Post-qualification, the following document(s) are required to be submitted within five (5) calendar days from receipt of notice from the BAC:</p> <ol style="list-style-type: none"> <li>1. Income Tax Returns for year 2018 (BIR Form 1701 or 1702).</li> <li>2. Latest Value Added Tax Returns (Forms 2550M and 2550Q) or Percentage Tax Returns (Form 2551M) for the last six (6) months before the deadline of Opening of Bids.</li> </ol> <p><i>The income tax and business tax returns stated above should have been filed through the Electronic Filing and Payment System (EFPS).</i></p> <ol style="list-style-type: none"> <li>3. Proof of completion of the single largest contract as identified in the Statement of Single Largest Contract, which shall be copy of any verifiable document(s) such as but not limited to the following: (a) Contract/s or Purchase Order/s; (b) corresponding Sales Invoice/s; (c) Official Receipt/Cash Receipt/Collection Receipt; and (d) Certificate of Satisfactory Completion.</li> <li>4. Submission of evidences<sup>1</sup> as proof of compliance with the bidder's actual offer, if applicable.</li> <li>5. <b>Updated PhilGEPS Certificate of Registration (Platinum Membership), in the event that the PhilGEPS Certificate and the annexes submitted during Opening of Bids is not updated.</b></li> <li>6. The bidder shall submit any of the following <b>documents</b> whichever is applicable: <ol style="list-style-type: none"> <li>6.1 If the bidder is a manufacturer, certify that the bidder manufactures the products/items or;</li> <li>6.2 If the bidder is an Exclusive/Authorized Distributor or Dealer of the products/items; <ol style="list-style-type: none"> <li>6.2.1 Certificate or Contract from the manufacturer must be provided as proof that the bidder is an Exclusive/Authorized Distributor or Dealer of the products/items or;</li> </ol> </li> <li>6.3 If the bidder is an agent of the exclusive distributor or dealer the following must be provided; <ol style="list-style-type: none"> <li>6.3.1 Certificate or Dealership Agreement by the manufacturer with the distributor or dealer.</li> </ol> </li> </ol> </li> </ol> <p><i>N.B. Documents submitted during post-qualification as part of post-qualification documents must be certified by the authorized representative to be true copy/ies from the original.</i></p>
------	--

<sup>1</sup> In the column "Bidder's Compliance", the bidder must state "comply" against each of the individual parameters of each specification corresponding to performance parameter of equipment offered. Statement of "comply" must be supported by evidence in a bidders bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's or distributor's un-amended sales literature, unconditional statements or specification and compliance issued by the manufacturer or distributor, samples, independent test data etc., as appropriate.

## Section VII. Technical Specifications

LOT NO. 1	:	<b>Supply and Delivery of Eye Screening Equipment</b>
ITEM NO.1	:	<b>Snellen Chart, Lighted</b>
QUANTITY	:	1,419 <u>units</u>

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE
<b>Snellen Chart, Lighted</b>	Brand and model:
Illuminated Snellen Eye Test Chart	
Height: 24 inches	
Width: 12 inches	
Thickness: 8 inches ( <b><u>maximum</u></b> )	
Voltage: 220V AC, 60Hz	
With <b><u>maximum of</u></b> five (5) bulbs/ <b><u>tubes</u></b>	
<b><u>With minimum total power consumption of</u></b> 25 watts	
Non-reflective matte finish	
<b><u>Printing on the Snellen Chart will be in Letters</u></b>	
<b><u>Can be read at a distance of 10 feet</u></b>	
Warranty: Must have at least one (1) year warranty upon delivery of goods.	
REQUIRED DOCUMENTS:  1. The bidder shall submit any of the following <b><u>documents</u></b> whichever is applicable:  1.1 If the bidder is a manufacturer, certify that the bidder manufactures the products/items or;  1.2 If the bidder is an Exclusive/Authorized Distributor or Dealer of the products/items;  1.2.1 Certificate or Contract from the	



<p>manufacturer must be provided as proof that the bidder is an Exclusive/Authorized Distributor or Dealer of the products/items or;</p> <p>1.3 If the bidder is an agent of the exclusive distributor or dealer the following must be provided;</p> <p>1.3.1 Certificate or Dealership Agreement by the manufacturer with the distributor or dealer.</p>	
<p><b>Packaging Instructions:</b></p> <p>Standard packaging of the manufacturer including product manual/instructional guide.</p>	
<p><b>Labeling Instructions:</b></p> <p>On each unit, the following should be legibly <b>imprinted or stickered</b> with non-removable or permanent sticker/label that is binding and with residue and tearing, if removed:</p> <p style="padding-left: 40px;">Philippine Government Property –Department of Health</p> <p style="padding-left: 40px;"><b>NOT FOR SALE</b></p> <p>On each corrugated carton, the following should be legibly <b>imprinted or stickered</b> with non-removable or permanent sticker/label that is binding and with residue and tearing, if removed:</p> <p style="padding-left: 40px;">Philippine Government Property – Department of Health</p> <p style="padding-left: 40px;"><b>NOT FOR SALE</b></p>	
<p><b>Product Recall/Replacement:</b></p> <p>The supplier must ensure the quality of the equipment and if there will be problems in the quality, the supplier will recall and replace the equipment distributed in the regions and health centers.</p>	

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found false either during evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company	Signature Over Printed Name of Authorized Representative	Date
-----------------	---	------

LOT NO. 1	:	<b>Supply and Delivery of Eye Screening Equipment</b>
ITEM NO.2	:	<b>Eye Occluder, Plastic</b>
QUANTITY	:	3,000 sets

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE
<b>Eye Occluder, Plastic</b>	Brand and model:
Set of 6 durable plastic occluder	
Completely covers the eye cupped for comfort	
Diameter: 4.5cm (minimum)	
Length: 15.6 cm (minimum)	
Color: preferably "Black"	
Material: should be latex-free	
Accessories: Carrying case with zipper	
Warranty: Must have at least one (1) year warranty upon delivery of goods.	
REQUIRED DOCUMENTS:  1. The bidder shall submit any of the following <b><u>documents</u></b> whichever is applicable:  1.1 If the bidder is a manufacturer, certify that the bidder manufactures the products/items or;  1.2 If the bidder is an Exclusive/Authorized Distributor or Dealer of the products/items;  1.2.1 Certificate or Contract from the manufacturer must be provided as proof that the bidder is an Exclusive/Authorized Distributor or Dealer of the products/items or;  1.3 If the bidder is an agent of the exclusive distributor or dealer the following must be provided;  1.3.1 Certificate or Dealership Agreement by the manufacturer with the distributor or	

dealer.	
<b>Packaging Instructions:</b> Standard packaging of the manufacturer including product manual/instructional guide.	
<b>Labeling Instructions:</b> On each unit, the following should be legibly <b>imprinted:</b>  Philippine Government Property – Department of Health  <b>NOT FOR SALE</b>  On each corrugated carton, the following should be legibly <b>imprinted or stickered</b> with non-removable or permanent sticker/label that is binding and with residue and tearing, if removed:  Philippine Government Property – Department of Health  <b>NOT FOR SALE</b>	
<b>Product Recall/Replacement:</b> The supplier must ensure the quality of the equipment and if there will be problems in the quality, the supplier will recall and replace the equipment distributed in the regions and health centers.	

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found false either during evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

_____ Name of Company	_____ Signature Over Printed Name of Authorized Representative	_____ Date
--------------------------	--	---------------

LOT NO.1	:	<b>Supply and Delivery of Eye Screening Equipment</b>
ITEM NO.3	:	<b>Pinhole Eye Occluder, Plastic</b>
QUANTITY	:	3,000 units

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE
<b>Pinhole Eye Occluder, Plastic</b>	Brand and model:
Handheld monocular occluder	
Single pinhole	
To cover either eye during an eye examination	
Diameter: 4.5 cm (minimum)	
Length of handle: 9 inches (maximum)	
Color: preferably "Black"	
Material: should be latex-free	
Accessories: Carrying case with zipper	
Warranty: Must have at least one (1) year warranty upon delivery of goods.	
<p><b>REQUIRED DOCUMENTS:</b></p> <p>1. The bidder shall submit any of the following <b><u>documents</u></b> whichever is applicable:</p> <p style="padding-left: 40px;">1.1 If the bidder is a manufacturer, certify that the bidder manufactures the products/items or;</p> <p style="padding-left: 40px;">1.2 If the bidder is an Exclusive/Authorized Distributor or Dealer of the products/items;</p> <p style="padding-left: 80px;">1.2.1 Certificate or Contract from the manufacturer must be provided as proof that the bidder is an Exclusive/Authorized Distributor or Dealer of the products/items or;</p> <p style="padding-left: 40px;">1.3 If the bidder is an agent of the exclusive distributor or dealer the following must be provided;</p>	

<p>1.3.1 Certificate or Dealership Agreement by the manufacturer with the distributor or dealer.</p>	
<p>Packaging Instructions: Standard packaging of the manufacturer including product manual/instructional guide.</p>	
<p>Labeling Instructions: On each unit, the following should be legibly <b>imprinted</b>:</p> <p style="padding-left: 40px;">Philippine Government Property –Department of Health</p> <p style="padding-left: 40px;"><b>NOT FOR SALE</b></p> <p>On each corrugated carton, the following should be legibly <b>imprinted or stickered</b> with non-removable or permanent sticker/label that is binding and with residue and tearing, if removed:</p> <p style="padding-left: 40px;">Philippine Government Property – Department of Health</p> <p style="padding-left: 40px;"><b>NOT FOR SALE</b></p>	
<p>Product Recall/Replacement: The supplier must ensure the quality of the equipment and if there will be problems in the quality, the supplier will recall and replace the equipment distributed in the regions and health centers.</p>	

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found false either during evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company	Signature Over Printed Name of Authorized Representative	Date
-----------------	---	------

LOT NO.1	:	<b>Supply and Delivery of Eye Screening Equipment</b>
ITEM NO.4	:	<b>Diagnostic Set with Ophthalmoscope, Fiber-Optic Oscope</b>
QUANTITY	:	2,100 sets

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE
<b>Diagnostic Set with Ophthalmoscope, Fiber-Optic Oscope</b>	Brand and model:
<b>Ophthalmoscope</b>	
<ul style="list-style-type: none"> <li>• Pocket type</li> <li>• Ophthalmoscope</li> <li>• Illumination/Direct Illumination: 18.5 Lumen <b>(minimum)</b></li> <li>• LED</li> <li>• 2.5 volts <b>(minimum)</b></li> <li>• <b>One (1) piece (head and handle)</b></li> </ul>	
<b>Fiber-Optic Oscope</b>	
<ul style="list-style-type: none"> <li>• High Performance Fibre Optics with 4mm diameter</li> <li>• 2.5V LED <b>or</b> Xenon &amp; Vacuum Illuminations</li> <li>• <b>One (1) piece</b> (head and handle)</li> <li>• 3x magnifying lens and wide viewing window</li> <li>• Connection via screw thread and dimmable light control system</li> <li>• <b>Five (5)</b> Disposable ear funnels <b>per</b> size: 2.5 mm, 3.0mm, 4.0mm</li> </ul>	
<b>Accessories:</b>	
<ul style="list-style-type: none"> <li>• Hard Storage</li> <li>• Accessory bumper kit</li> <li>• Onyx <b>color</b></li> <li>• Battery included (AA batteries)</li> </ul>	
Warranty: Must have at least five (5) years warranty upon delivery of goods. <b><u>Warranty of the unit does not include the bulb.</u></b>	
<b>REQUIRED DOCUMENTS:</b>	
<ol style="list-style-type: none"> <li>1. The bidder shall submit any of the following <b><u>documents</u></b> whichever is applicable: <ol style="list-style-type: none"> <li>1.1 If the bidder is a manufacturer, certify that the bidder manufactures the products/items or;</li> <li>1.2 If the bidder is an Exclusive/Authorized Distributor or Dealer of the products/items; <ol style="list-style-type: none"> <li>1.2.1 Certificate or Contract from the</li> </ol> </li> </ol> </li> </ol>	

<p>manufacturer must be provided as proof that the bidder is an Exclusive/Authorized Distributor or Dealer of the products/items or;</p> <p>1.3 If the bidder is an agent of the exclusive distributor or dealer the following must be provided;</p> <p>1.3.1 Certificate or Dealership Agreement by the manufacturer with the distributor or dealer.</p>	
<p>Packaging Instructions:</p> <p>Standard packaging of the manufacturer including product manual/instructional guide.</p>	
<p>Labeling Instructions:</p> <p>On each unit, the following should be legibly <b>imprinted or stickered</b> with non-removable or permanent sticker/label that is binding and with residue and tearing, if removed:</p> <p style="padding-left: 40px;">Philippine Government Property – Department of Health</p> <p style="padding-left: 40px;"><b>NOT FOR SALE</b></p> <p>On each corrugated carton, the following should be legibly <b>imprinted or stickered</b> with non-removable or permanent sticker/label that is binding and with residue and tearing, if removed:</p> <p style="padding-left: 40px;">Philippine Government Property – Department of Health</p> <p style="padding-left: 40px;"><b>NOT FOR SALE</b></p>	
<p>Product Recall/Replacement:</p> <p>The supplier must ensure the quality of the equipment and if there will be problems in the quality, the supplier will recall and replace the equipment distributed in the regions and health centers.</p>	

I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found false either during evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company	Signature Over Printed Name of Authorized Representative	Date
-----------------	---	------