



**Bid Bulletin No. 1
 November 28, 2019**

PUBLIC BIDDING NO. 19-326-5

**SUPPLY AND DELIVERY OF PLASMA FREEZERS FOR THE DEPARTMENT OF HEALTH
 - NATIONAL VOLUNTARY BLOOD SERVICES PROGRAM (DOH-NVBSP)**

Issued pursuant to Sec. 22.5 of the IRR of R.A. 9184 to clarify and/or amend certain provisions in the Bidding Documents issued for this project, considering the issues raised and clarifications made by prospective bidders during the **Pre-Bid Conference** held on **26 November 2019**, likewise, respond to bidders' written queries received within the prescriptive period for filing.

A. AMENDMENTS

REFERENCE	BASES FOR AMENDMENT
Section VII. TECHNICAL SPECIFICATIONS For Lot 1 Page 72 xxx AS AMENDED: Operating temperature: Temperature can be set to -18° C to -30°C (Colder temperature range is acceptable) with setting accuracy of <u>no more than</u> +/- 1°C whatever the load. xxx	<i>As clarified by the End-User.</i>

B. CLARIFICATIONS

The Department of Budget and Management - Procurement Service, Bids and Awards Committee hereby clarifies prospective bidders' concerns and queries:

ITEM	CONCERN	REFERENCE	CLARIFICATION/ RESOLUTION
1. LABTRADERS, INC.			
1	Request to consider a Manufacturer's certification that the product to be delivered is complying with the requirement. Released brochure for the product is 230 V/50 Hz.	Section VII. Technical Specifications LOT 1 Pages 73 xxx Power Supply Input: 220-240 VAC, 60Hz xxx	<i>As clarified by the End-User, original requirement is retained.</i>



All other portions of the Bidding Documents affected by these amendments shall be made to conform to the same.
Amendments/inclusions/clarifications made herein shall be considered an integral part of the Bidding Documents.

(SGD)
WEBSTER M. LAUREÑANA
Chairperson, Bids and Awards Committee V



TECHNICAL SPECIFICATIONS

LOT NO. 1 : **Upright Plasma Freezer**, at least 550 plasma bags
QUANTITY : **5 units**
APPROVED BUDGET FOR THE CONTRACT : **P 4,750,000.00**

AGENCY SPECIFICATIONS	BIDDER'S STATEMENT OF COMPLIANCE	REFERENCE
ITEM DESCRIPTION:	Brand and Model:	
Power Supply Input: 220 – 240 VAC, 60Hz		
Capacity: At least 550 plasma bags or its equivalent volume in liters		
Standard Features:		
<ul style="list-style-type: none"> • Plasma Freezers Equipment meets internationally accepted electrical safety specifications such as that IEC. 		
<ul style="list-style-type: none"> • Type of equipment: Compression freezer or its equivalent and utilizing CFC free refrigerant gas or its acceptable alternative 		
<ul style="list-style-type: none"> • Construction: Internal: Stainless steel or its equivalent 		
<ul style="list-style-type: none"> • External: corrosion resistant or its equivalent 		
<ul style="list-style-type: none"> • Upright type 		
<ul style="list-style-type: none"> • Solid door with roll out type drawers/shelves 		
<ul style="list-style-type: none"> • Internal temperature Control: Electronic temperature control 		
<ul style="list-style-type: none"> • Operating temperature: Temperature can be set to -18° C to -30°C (Colder temperature range is acceptable) with setting accuracy of no more than +/- 1°C whatever the load. 		
<ul style="list-style-type: none"> • Fan air cooling, automatic defrost within safe temperature control 		
<ul style="list-style-type: none"> • External ambient temperature: Performs in an ambient temperature +15°C to +32°C (Wider temperature range is acceptable) 		
<ul style="list-style-type: none"> • Temperature Monitoring: Digital temperature (LED) display with at least 0.1° C graduation 		
<ul style="list-style-type: none"> • Temperature recording device 		
<ul style="list-style-type: none"> • Visual and audible alarm system indicating unsafe temperatures 		



<ul style="list-style-type: none"> Battery backup for alarm and temperature recording device 		
<ul style="list-style-type: none"> Facility for remote alarm contact 		
Standard Accessories: <ul style="list-style-type: none"> Automatic Voltage Regulator (compatible to the equipment power requirement) 		
Miscellaneous: <ul style="list-style-type: none"> The unit being offered must be brand new and not a discontinued model. 		
<ul style="list-style-type: none"> Provide user's Manual in English Language <ul style="list-style-type: none"> Operations Manual 2 copies Service Manual 2 copies 		
<ul style="list-style-type: none"> The supplier has a capability for corrective and preventive maintenance of the unit 		
<ul style="list-style-type: none"> The parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period. 		
<ul style="list-style-type: none"> The supplier will be responsible for the notification, transportation to site, delivery, installation and testing on the site (hospital / health facility) and expenses for such will be on the account of the supplier. 		
<ul style="list-style-type: none"> Training Program for users and maintenance personnel of the hospital on equipment operation 		
<ul style="list-style-type: none"> Warranty of one (1) year for parts and two (2) years on service, upon delivery, inspection and acceptance. 		
<ul style="list-style-type: none"> Conformity with the appropriate ISO 13485 and/or CE Certificate 		
<ul style="list-style-type: none"> Availability of spare parts for the next 5 years after the warranty period. 		
<ul style="list-style-type: none"> Lifetime of product is at least 7 years. 		
<ul style="list-style-type: none"> The brand has been in the local or international market for at least ten (10) years. 		
<ul style="list-style-type: none"> Product Demonstration in Metro Manila 		
Labeling Instructions: <ul style="list-style-type: none"> Each Upright Plasma Freezer must have non-removable/ permanent markings "Philippine Government Property-Department of Health-Not For Sale", 1 inch in size on front surface. 		



I hereby certify that the statement of compliance to the foregoing technical specifications are true and correct, otherwise, if found to be false either during bid evaluation or post-qualification, the same shall give rise to automatic disqualification of our bid.

Name of Company

Signature Over Printed Name of
Authorized Representative

Date